

ADDENDUM-III

Nov. 30, 2013

Ref. Tender No. BGL/234/2013-14 dated 11.11.2013

Sub: Tender for Procurement of Mass Flow Meters (MFM) for CNG STATIONS Hyderabad

This is for general information to all bidders.


Sr No	Section Detail	Page No.	Point No.	Tender Description	Shall be read as
	Technical			As per Addendum -I	
3	II of II	16 of 22	1.10-b	Certificate of radiography/ x-ray for any welded joint to be submitted if available	Certificate of Radiography / X-ray for any welded joint
14	vol II of II	19 of 22	4.1	Mass Flow Meter Details: Sensor (12) Enclosure -- WP to IP 65 & Transmitter (21) Enclosure -- FLAMEPROOF + WP (IP65) & Wherever IP65 is mentioned elsewhere in the tender	Mass Flow Meter Details: Sensor (12) Enclosure -- WP to IP65 of better Ingress Protection & Transmitter (21) Enclosure -- FLAMEPROOF + WP (IP67)

Bidder shall take in to account the above while submitting their bid.

The Addendum – I shall form the integral part of the tender document.

All other Terms and Conditions of the tender document shall remain unaltered

**For and on Behalf of
M/s. Bhagyanagar Gas Limited**


**Bonny K Joseph
Dy. Manager (C&P)**